

**Technical Specifications:**

- Insulation Measurement Range: 4 ~ 4000M Ω (minimum)
- Accuracy: \pm (5% + 5d)
- DCV and ACV Measurement: 600V (minimum)
- Accuracy: \pm (1.5% + 2d)
- Display: Min. 25mm LCD
- Max Indication: 6100 (V) and 4200 (Ω & M Ω)
- Sampling Time: 0.3 ~ 0.5 sec
- Built-in over input and low battery indication facility
- Continuity beeper and data hold function
- Automatic zero adjust and circuit discharge facility
- Overload protection for each range
- LSI circuit to provide high reliability
- Accessories: Test lead, Alligators, Battery and Instruction manual
- Warranty: 1(One) year with services
- Documents: Manufacturer's Authorization Letter with contact details should be

Note: Specifications are subject to change, Photos shown above are Indicative, Actual Product can Vary.



Export Sales: +91-9829132777
India Sales: +91-9588842361



IT-2013, Ramchandrapura Industrial Area,
Sitapura Extension, Jaipur-302022, India.



info@tesca.in
www.tescaglobal.com

submitted

- ISO Certificate and Catalogue should be submitted



Note: Specifications are subject to change, Photos shown above are Indicative, Actual Product can Vary.



Export Sales: +91-9829132777
India Sales: +91-9588842361



IT-2013, Ramchandrapura Industrial Area,
Sitapura Extension, Jaipur-302022, India.



info@tesca.in
www.tescaglobal.com