



Apparatus consist a mild steel angle of size 1" x 1" x 1/8" or an equivalent metric units of length 80cm 80cm is as a cantilever beam. The beam should be fixed at one end such that the rotation of 45° intervals are given and clamped such that the principal axis of its cross-section may be inclined at any angle with the horizontal and vertical planes. Also arrangement will be provided to apply vertical load at the free end of the cantilever and to measure horizontal and vertical deflection of the free end. A dial gauge with 25mm travel (with magnetic base) will be supplied with the apparatus will be supplied complete with a supporting stand and a set of weights.

Note: Specifications are subject to change.

Tesca Technologies Pvt. Ltd.

IT-2013, Ramchandrapura Industrial Area, Sitapura Extension,
Near Bombay Hospital, Vidhani Circle, Jaipur-302022, Rajasthan, India,
Tel: +91-141-2771791 / 2771792; Email: info@tesca.in, tesca.technologies@gmail.com
Website: www.tesca.in