



55705 Experimental Set Up has been designed specifically to determine R the resistance of a Galvanometer by Kelvin's method using P.O. Box. The set up is absolutely self contained and requires no other apparatus. Practical experience on this set up carries great educative value for Science and Engineering Students.

OBJECT

01 To determine the resistance of a Galvanometer by Kelvin's method using P.O.Box.

FEATURES

The Set up consists of the following :

- 01 Post Office Box (Dial Type) OMEGATYPE POB-180.
- 02 Galvanometer 50-0-50 on bakelite Stand OMEGATYPE MO-65.
- 03 Leclanche Cell or substitute Cell Eliminator OMEGATYPE CE-1V5.
- 04 Decade Resistance Box, One dial in step of 100 ohms, total 1000 ohms OMEGATYPE DRBC-115R.
- 05 Weight : 4 Kg. (Approx.)
- 06 Adequate no. of connecting wires, 100cm long.
- 07 Strongly supported by detailed Operating Instructions, giving details of Object, Theory, Design procedures, Report Suggestions and Book References.

Note: Specifications are subject to change.

Tesca Technologies Pvt. Ltd.

IT-2013, Ramchandrapura Industrial Area, Sitapura Extension,
Near Bombay Hospital, Vidhani Circle, Jaipur-302022, Rajasthan, India,
Tel: +91-141-2771791 / 2771792; Email: info@tesca.in, tesca.technologies@gmail.com
Website: www.tesca.in

