

Order Code - 58012



#### TECHNICAL SPECIFICATION

- The valid temperature range is from 15°C to 65°C
- The valid sampling time range is from 30s to 10min,
- The valid peristaltic pump speed range is from 1 to 12
- The sampling speed is about 50ml/min.
- Power supply is 220±22V@50±1Hz or 110±11V@60±1Hz.

#### ACCESSORIES

- Control Unit.
- Cell holder with Peltier System. (It's already pre-loaded into the compartment of the spectrophotometer).
- Control Cable (to connect the Control Unit with the Cell holder with Peltier System).
- Peristaltic pump pipe. (It's already pre-loaded into the pump valve of the Control Unit)
- Power cord.



Note: Specifications are subject to change.

#### **Tesca Technologies Pvt. Ltd.**

IT-2013, Ramchandrapura Industrial Area, Sitapura Extension,  
Near Bombay Hospital, Vidhani Circle, Jaipur-302022, Rajasthan, India,  
Tel: +91-141-2771791 / 2771792; Email: info@tesca.in, tesca.technologies@gmail.com  
Website: www.tescaglobal.com