



Applications

The water bath in general will fulfill the requirements of IS:4031 (Part 11) Method of Physical tests for hydraulic cement (Determination of Density),

Technical Details

- Microprocessor based digital temperature controller system.
- Capable of maintaining temperature within + 0.1 °C accuracy
- Drainage System
- Stand for each Le-Chatlier Flask cap. 250ml (8 Nos.)
- It will be attached with cooling & heating arrangement system.
- It will be fully covered.
- Water bath body will be insulated.
- Control panel will be at side of the bath.
- Water circulation/stirrer will be provided to maintain uniform temperature of water.
- Transparent (toughened glass) sheet will be provided in front and back position.

Technical Specification

Volume of the bath	: 48 liter
Material of construction	: Inner S.S. 304 and Outer M.S. with Powder Coating
Working size	: 58 x 25 x 30 cm
Outer Size	: 145 x 48 x 64 cm

Note: Specifications are subject to change.