



Compliance Standards:

ISO 10545 (Part-7), BIS 13630 (Part-XI)

Details of the apparatus:

- The apparatus consists of a viewing box equipped with fluorescent lighting
- Fluorescent light offers color temperature of 6000 K to 6500 K
- The lightings are placed vertically above the surface to be observed providing 300 lux illuminance.
- The dimensions of the box is 61 x 61 x 61 cm
- The inner faces are painted a neutral grey.
- Light source is screened to avoid direct viewing.

Overall Dimensions in Inches (L x W x H): $24 \times 24 \times 30$

Note: Specifications are subject to change.

Near Bombay Hospital, Vidhani Circle, Jaipur-302022, Rajasthan, India,

Tel: +91-9829132777; Email: info@tesca.in, tesca.technologies@gmail.com

Website: www.tescaglobal.com