

**Technical Specifications :**

- Complies to AASHTO T 308-99 and ASTM D6307-98
- Used for the quantitative determination of bitumen in hot-mixed paving mixtures and pavement samples.
- The bitumen content is calculated by difference from the weight of extracted aggregates, moisture content and ash from an aliquot part of the extract.
- The apparatus consists of two wire mesh cones with interlocking frames, a cylindrical glass jar and a water condenser with inlet/outlet tubes.

Note: Specifications are subject to change, Photos shown above are Indicative, Actual Product can Vary.



Export Sales: +91-9829132777
India Sales: +91-9588842361



IT-2013, Ramchandrapura Industrial Area,
Sitapura Extension, Jaipur-302022, India.



info@tesca.in
www.tescaglobal.com