

Order Code : 20213501.1.6
Name : Reynold's Apparatus



The Set-up is designed to verify Reynold's Apparatus experimentally. The Apparatus consists of a glass tube which one end having bell mouth entrance connected to a water tank. At the other end of the glass tube a cock is provided to vary the rate of flow. Flow rate of water can be measured with the help of Measuring Cylinder and Stop Watch, supplied with the set-up. A capillary tube is introduced centrally in the bell mouth. To this tube dye is fed from a small container, placed at the top of Constant head Tank, through polythene tubing.

EXPERIMENTS:

- ∞ To study different types of flow.
- ∞ To determine the Reynold's Number and hence the type of flow either laminar or turbulent.

UTILITIES REQUIRED:

- ∞ Water Supply.
- ∞ Drain.

TECHNICAL SPECIFICATION:

- ∞ Tube : Material Borosilicate Glass
- ∞ Dye Vessel : Material SS, Suitable Capacity
- ∞ Capillary Tube : Material Copper/Stainless Steel
- ∞ Constant Head Water Tank : Capacity 40 Liters.
- ∞ Water Circulation : FHP Pump
- ∞ Flow Measurement : Using Measuring Cylinder
- ∞ Sump Tank : Capacity 60 Liters.
- ∞ Stop Watch : Capacity 60 Liters. Electronic
- ∞ Control Panel Comprises of : Standard make, On/Off Switch, Mains Indicator, etc.
- ∞ The Whole set-up is well designed and arranged in a good quality painted structure.

Note: Specifications are subject to change, Photos shown above are Indicative, Actual Product can Vary.

TESCA TECHNOLOGIES PVT. LTD.

IT-2013, Ramchandrapura Industrial Area, Sitapura Extension, Jaipur-302029, Rajasthan, India.
Ph/ Fax: 91-141-2771791, 2771792; Email: info@tesca.in, tesca.technologies@gmail.com
Website: www.tesca.in