



### Key Features

- DETERMINATION OF THE REFRACTIVE INDEX OF HOLLOW PRISM USING SPECTROMETER AND MERCURY LIGHT
- **ABOUT SPECTROMETER**
- A Basic measuring instrument for quantitative spectroscopy experiments, this instrument is mounted on a stable cast iron base with attached collimator and a rotating telescope platform and graduated circle.
- A 2.4cm. Diameter achromatic optical system with a 10x Ramsden eyepiece & cross line graticule.
- An 8 cm. Diameter prism/grating Table, fully adjustable with prism & grating holders and engraved ring pattern. Protected Circle 15.0 cm with S.S. Scale.
- Vernier reading 1 minute.
- Supplied in a wooden case with 2" Brass spirit level.

### REQUIRED ACCESSORIES

#### **Apparatus Supply:**

- Spectrometer
- EDF Prism 32x32
- Prism 50x50
- Spirit Level 2"
- Working Manual

#### **Accessories (Light Source):**

- Mercury Vapour Lamp Assembly
- Mercury Vapour Lamp 80 Watts
- Mercury Vapour Lamp Box
- Mercury Vapour lamp Transformer 80 Watts

Note: Specifications are subject to change, Photos shown above are Indicative, Actual Product can Vary.



Export Sales: +91-9829132777  
India Sales: +91-9588842361



IT-2013, Ramchandrapura Industrial Area,  
Sitapura Extension, Jaipur-302022, India.



info@tesca.in  
www.tescaglobal.com