

**Technical Specifications :**

- The Aggregate Crushing Value (ACV) test set provides a relative measure of the resistance of an aggregate to crushing under a gradually applied compressive load.
- All parts of the apparatus are powder coated or galvanized steel, heat treated and ground before manufacturing to make it durable and reliable.
- M.S. Cylindrical Container 150mm +/-0.5mm dia x 130mm to 140mm high with base plate 200 to 230mm/sqr x 6mm thick.
- A Plunger of 148mm +/-0.5mm dia x 100 to 115mm high.
- Supplied complete with Tamping Rod, 16mm dia x 600mm long, both ends rounded, 1 no. Metal Measure 115+/- 0.5mm dia x 180 +/- 0.5mm high.

*Note: Specifications are subject to change, Photos shown above are Indicative, Actual Product can Vary.*



Export Sales: +91-9829132777  
India Sales: +91-9588842361



IT-2013, Ramchandrapura Industrial Area,  
Sitapura Extension, Jaipur-302022, India.



info@tesca.in  
www.tescaglobal.com