



Order Code : 20213547.5.5

Name : Ultrasound diffraction apparatus



Specifications:

- Input : 230 volts AC 50c/s
- Frequency : 2.5-6.5Mc/s calibrated for different set of crystal Freq. & is Continuously variable R.F.
- Output : The output is brought out on the terminals. Built in a nicely Polished box with on/off switch jewel light etc.
- Accessories Sodium Lamp & TXFR : Complete with 35 Watt TXFR & woodencase.
- Ceramic Crystal : The Crystal is provided with the handle.
- Tank : Optically true glass tank is provided for keeping the Liquid.
- Liquid : Kerosene Oil is provided in a plastic bottle

Note: Specifications are subject to change, Photos shown above are Indicative, Actual Product can Vary.

TESCA TECHNOLOGIES PVT. LTD.

IT-2013, Ramchandrapura Industrial Area, Sitapura Extension, Jaipur-302029, Rajasthan, India.
Ph/ Fax: 91-141-2771791, 2771792; Email: info@tesca.in, tesca.technologies@gmail.com
Website: www.tesca.in